

(D)

PLOT SCALE: 1/4=1

FIG. SCALE:1



1/8 THK. LOW CARBON STEEL SHEET



PREPARATION
AS NEEDED
TO PRODUCE
SEALED CORNER

CORNER SEALED WELD

SHEET BLANK @ corner

SIDING PREPARATION TYPICAL DETAIL

NOTES

1. LEAD PANELS SHALL BE CONSTRUCTED FROM SINGLE UNPIERCED LEET. FREE OF OXIDE INCLUSIONS, FISSURES, VOIDS AND ALL OTHER DETRIMENTAL DEFECTS. MATERIAL SHALL LEAD GRADE "C" TO COMPLY WITH FEDERAL SPECIFICATION 48020. CABINET SHALL MEET ALL APPLICABLE FEDERAL, STATE AND LOCAL RADIATION GUIDE LINES: A.S. HANDBOOK 111 AND 114, A.S. N43.2 AND ORDER 5480.11.
2. FINISH - PAINT EXTERIOR SURFACES, EXCEPT AS NOTED, WITH PRATT & LAMBERT #2127 (SNO-FLAKE WHITE, SEMI-GLOSS, EPOXY BASE) OR EQUIVALENT, SURFACE PREPARATION PER A.N.L. SPECIFICATION #4105-000 24.
3. DIMENSIONS IN BRACKETS [] ARE MILLIMETERS AND ARE FOR REFERENCE ONLY.
4. SOME HIDDEN LINES MAY BE OMITTED FOR CLARITY.

[illegible]

U6 ID SHIELDED CABINET
SLITS L2
ROOF PANEL